

NEIDL Agent Incident Reporting Summary April 2018 - June 2018, Q2

Date of Incident	Incident Type / Agent Involved	BSL	Transmissible Person to Person	Description	Reportable Incident	Reportable of Clinical Illness	Agency Reported To	Comments / Corrective Actions
6/13/18	Breach with BSL-4 outer suit zipper	BSL4	No	A facilities staff person noticed his left leg scrub pants were wet while in the decontamination shower exiting BSL4 containment. He was further evaluated in ROHP.	Yes	No	BPHC	<p>During this incident the lab was not hot (no RG4 viruses present in the space). This incident was discussed with the Facility manager and the decision was made to avoid moving BSC while the lab is hot. We will try very hard to never resource to that option in the future, however, if we have to do it again, the activity will be closely monitored/supervised by EHS.</p>